


GW501516 HPLC

Client :SARMS4LESS
Test time :05/07/2021
Product Name :GW501516
Analyst :GSL
Lot No. :20210507
Column: :GSA46250BI017013, C18, 4.6*250mm, 5um
Solvent A :100% Acetonitrile
Solvent B :0.1% Trifluoroacetic in 100% Water
Gradient :
 A B
 0.0min 5% 95%
 20.0min 95% 5%
 20.1min 100% 0%
 30.0min Stop
Flow Rate :1.0ml/min
Wavelength :220nm
Temperature :40°C
Concentration :1mg/ml
Sample Volume :6ul



No.	Stay time	Concentration	Peak area
1	13.743	0.5221	25958
2	16.331	98.81	4912740
3	16.953	0.6616	32894
Total		100	4971592

WWW.SARMS4LESS.COM