SR9009 HPLC Report

Client	:SARMS4LESS
Test time	:05/11/2021
Product Name	:SR9009
Analyst	:LJ
Lot No.	:20210511
Column:	:GSA46250BIO17013,C18，4.6*250mm,5um
Solvent A	:100% Acetonitrile
Solvent B	:0.1% Trifluoroacetic in 100% Water
Gradient	: A B
	0.0min 60% 40%
	25.0min 80% 20%
	25.1min 100% 0%
	30.0min Stop
Flow Rate	:1.0ml/min
Wavelength	:220nm
Temperature	:37℃
Concentration	:1mg/ml
Sample Volume	:10ul



WWW.SARMS4LESS.COM

Number	Stay time	Concentration	Peak area	Peak height
1	10.352	0.103078	56	24
2	10.940	99.0	18159581	1683703
Total		100	18159637	1683727